

SAN ANTONIO WATER SYSTEM PURCHASING DEPARTMENT

Issued By: Sandra Rios
BID NO.: 19-19122

Date Issued: October 2, 2019

FORMAL INVITATION FOR BIDS
FORMAL THE ONE TIME PURCHASE OF A DISCRETE ANALYZER
ADDENDUM 2

Sealed bids addressed to the Purchasing Director, San Antonio Water System, 2800 US Hwy 281 North, Administration Bldg., 5th Floor, San Antonio, TX 78212 will be received until **3:00 p.m. October 7, 2019** and then publicly opened and read aloud for furnishing materials or services as described herein below,

The San Antonio Water System Purchasing Department is willing to assist any bidder(s) in the interpretation of bid provisions or explanation of how bid forms are to be completed. Assistance may be received by visiting the Purchasing Office in the SAWS Main Office, 2800 US Hwy 281 North, San Antonio, TX 78212, or by calling (210) 233-3819.

This invitation includes the following:

Invitation for Bids
Terms and Conditions of Invitation for Bids

Specifications and General Requirements
Price Schedule

The undersigned, by his/her signature, represents that he/she is authorized to bind the Bidder to fully comply with the Specifications and General Requirements for the amount(s) shown on the accompanying bid sheet(s). By signing below, Bidder has read the entire document and agreed to the terms therein.

Signer's Name: _____ Firm Name: _____
(Please Print or Type)

Address: _____

Signature of Person Authorized to Sign Bid _____ City, State, Zip Code: _____

Email Address: _____ Telephone No.: _____

Fax No.: _____

Please complete the following:

Prompt Payment Discount: _____% _____days. (If no discount is offered, Net 30 will apply.)

Please check the following blanks which apply to your company:

Ownership of firm (51% or more):

Non-minority Hispanic African-American Other Minority (specify) _____

Female Owned Handicapped Owned Small Business (less than \$1 million annual receipts or 100 employees)

Indicate Status: Partnership Corporation Sole Proprietorship Other (specify) _____

Tax Identification Number: _____

To report suspected ethics violations impacting the San Antonio Water System, please call 1-800-687-1918.

***** This **Addendum 2** is issued for Bid 19-19122.

1. Answer question(s).

IT IS **NOT NECESSARY** TO RETURN THIS ADDENDUM 2 AS PART OF YOUR BID SUBMISSION

.....

1. Is the SEAL AQ400 discrete analyzer pre-approved to offer a bid response?

Response: The instrument will meet the requirements IF the options ('110 volt' and 'optional integrated cadmium coil reduction for nitrate/nitrite determination') in brochure are included in bid. All Hardware Specifications requested within solicitation are a requirement.

2. Does the warranty and service support for the 1st year begin one year after installation or at the time of installation?

Response: At the time of installation.

IT IS **NOT NECESSARY** TO RETURN THIS ADDENDUM 2 AS PART OF YOUR BID SUBMISSION

All other terms and conditions of the original bid remain unchanged.