

SAN ANTONIO WATER SYSTEM PURCHASING DEPARTMENT

Issued By: Angeline C. Peralez
BID NO.: 12 - 5032

Date Issued: July 23, 2012

**FORMAL INVITATION FOR BIDS
ANNUAL CONTRACT FOR CALIBRATION GAS CYLINDERS
ADDENDUM 1**

Sealed bids addressed to the Purchasing Director, San Antonio Water System, 2800 US Hwy 281 North, P.O. Box 2449, San Antonio, TX 78298-2449 will be received until **3:00 p.m., July 31, 2012** and then publicly opened and read aloud for furnishing materials or services as described received herein below,

The San Antonio Water System Purchasing Department is willing to assist any bidder(s) in the interpretation of bid provisions or explanation of how bid forms are to be completed. Assistance may be received by visiting the Purchasing Office in the SAWS Main Office, 2800 US Hwy 281 North, San Antonio, TX 78212, or by calling (210) 233-3819.

This invitation includes the following:

Invitation for Bids
Terms and Conditions of Invitation for Bids

Specifications and General Requirements
Price Schedule

The undersigned, by his/her signature, represents that he/she is authorized to bind the Bidder to fully comply with the Specifications and General Requirements for the amount(s) shown on the accompanying bid sheet(s). By signing below, Bidder has read the entire document and agreed to the terms therein.

Signer's Name: _____ Firm Name: _____
(Please Print or Type)

Address: _____

Signature of Person Authorized to Sign Bid _____ City, State, Zip Code: _____

Email Address: _____ Telephone No.: _____

Fax No.: _____

Please complete the following:

Prompt Payment Discount: _____% _____ days. (If no discount is offered, Net 30 will apply.)

Please check the following blanks which apply to your company:

Ownership of firm (51% or more):

___ Non-minority ___ Hispanic ___ African-American ___ Other Minority (specify) _____

___ Female Owned ___ Handicapped Owned ___ Small Business (less than \$1 million annual receipts or 100 employees)

Indicate Status: ___ Partnership ___ Corporation ___ Sole Proprietorship ___ Other (specify) _____

Tax Identification Number: _____

***** This Addendum 1 is issued to clarify the following on the bid mailed to you on July 11, 2012.

I. Questions and Clarifications.

II. Updated Price Schedule

I. Questions

Question 1: Will you accept substitution products for the 34 Liter, like 29 Liter?

Answer: No, 29 liter cylinder run out to fast we have too many pieces of equipment that we calibrate.

Question 2: What size cylinder are you asking for on the last line item 25 H2S, 100 CO, 2.5% CH4, 18% O2, N2 balance?

Answer: 58 Liter cylinder

Question 3: Please clarify this concentration: 200 ppm CO, 2.5% (50% LEL) CH4, 18.5% volume, Air balance N2 Humidified. Is the 18.5% Oxygen.

Answer: The concentration is 100 PPM on CO not 200 PPM and the Oxygen is 20.9% not 18.5%

Question 4: On the same concentration above, please clarify nitrogen Humidified.

Answer: Dry

II. Updated Price Schedule

The most updated Price Schedule is attached to this Addendum 1 to replace the Price Schedule on bid package mailed to you on July 11, 2012.

PRICE SCHEDULE

<u>Description</u>	<u>Quantity</u>	<u>Price/Unit</u>	<u>Extended Price</u>
<u>103 L Gas Cylinders 2.5% Methane</u> (50% LEL) Part No. 077-0037 Air Balance 4888 NIC II LEL Combustible Gas Detection Transmitter	7	\$ _____/cyl	\$ _____
<u>103 L Gas Cylinders</u> Oxygen 16.0% Part No. 077-0039 Nitrogen Balance Freedom 5000 Universal Analog Toxic Gas Transmitter	7	\$ _____/cyl	\$ _____

<u>58 L Gas Cylinders</u>	7	\$ _____/cyl	\$ _____
Chlorine 5 PPM			
Part No. 077-0152			
Nitrogen Balance Freedom 5000 Universal Analog Toxic Gas Transmitter			
<u>58 L Gas Cylinders</u>	7	\$ _____/cyl	\$ _____
Sulfur Dioxide 5 PPM			
Part No. 1810 2222			
Nitrogen Balance T-80 Portable Gas Detector			
<u>58 L Gas Cylinders</u>	7	\$ _____/cyl	\$ _____
Chlorine 10 PPM			
Part No. 1810 1758			
Nitrogen Balance T-80 Portable Gas Detector			
<u>58 L Gas Cylinders</u>	7	\$ _____/cyl	\$ _____
Sulfur Dioxide 10 PPM			
Part No. 009824-39			
Nitrogen Balance Sensidyne Gas Detectors			
<u>58 L Gas Cylinders</u>	7	\$ _____/cyl	\$ _____
Chlorine 5 PPM			
Part No. 0098 24-34			
Nitrogen Balance Sensidyne Gas Detectors			
<u>103 L Gas Cylinders</u>	7	\$ _____/cyl	\$ _____
Hydrogen Sulfide 25 PPM			
Carbon Monoxide 50 PPM			
Methane 1.62% (32.0% LEL) Oxygen 19.0%			
Part No. 077-0352			
Nitrogen Balance Scout Portable Detectors			
<u>103 L Gas Cylinder</u>	7	\$ _____/cyl	\$ _____
Air Zero Grade			
Part No. 009824-12			
THC<1 PPM Sensidyne Gas Detectors			
<u>34 L Gas Cylinders</u>	7	\$ _____/cyl	\$ _____
Sulfur Dioxide 10 PPM			
Part No. 077-0276			
Air Balance Scout Portable Detectors			
<u>58 L Gas Cylinders</u>	7	\$ _____/cyl	\$ _____
Hydrogen Sulfide 10 PPM			
Part No. 502103J (Scott Part)			
Nitrogen Balance Interscan Portable Detector 34 L Gas Cylinders			
<u>58 L Gas Cylinders</u>	7	\$ _____/cyl	\$ _____
Hydrogen Sulfide 25.0 PPM			
Part No. 430601J (Scott Part)			
Nitrogen Balance Interscan Portable Detector			

<u>34 L Gas Cylinders</u>	7	\$ _____/cyl	\$ _____
Hydrogen Sulfide 25 PPM			
Carbon Monoxide 50 PPM			
2.5 % Methane 50% LEL			
Part No. GFV243			
Air Balance Lumidor Portable Detector			

<u>AGM 502 calibration gas</u>	170	\$ _____/cyl	\$ _____
Part no. 7802-016			
GfG Instrumentation (no substitution)			
34 liters 100ppm CO			
2.5% Vol			
50% LEL			
CH4 20.9% Volume Air			
Balance nitrogen, dry			

<u>Multi-Gas for Micro 5</u>	7	\$ _____/cyl	\$ _____
H2S 25ppm			
58 Liter			
Carbon Monoxide 100ppm			
Methane 50% LEL			
Oxygen 18% Nitrogen Balanced			

GRAND TOTAL \$

IT IS NECESSARY TO RETURN THIS ADDENDUM 1 TOGETHER WITH YOUR BID.