

SAN ANTONIO WATER SYSTEM PURCHASING DEPARTMENT

Issued By: Sandra Rios
BID NO.: 19-19112A

Date Issued: September 23, 2019

**FORMAL INVITATION FOR BIDS FOR THE ONE TIME PURCHASE OF AUTOMATED
BIOCHEMICAL OXYGEN DEMAND (BOD) ANALYZER
ADDENDUM 2**

Sealed bids addressed to the Purchasing Director, San Antonio Water System, 2800 US Hwy 281 North, Administration Bldg., 5th Floor, San Antonio, TX 78212 will be received until **3:00 p.m. September 30, 2019** and then publicly opened and read aloud for furnishing materials or services as described herein below,

The San Antonio Water System Purchasing Department is willing to assist any bidder(s) in the interpretation of bid provisions or explanation of how bid forms are to be completed. Assistance may be received by visiting the Purchasing Office in the SAWS Main Office, 2800 US Hwy 281 North, San Antonio, TX 78212, or by calling (210) 233-3819.

This invitation includes the following:

Invitation for Bids
Terms and Conditions of Invitation for Bids

Specifications and General Requirements
Price Schedule

The undersigned, by his/her signature, represents that he/she is authorized to bind the Bidder to fully comply with the Specifications and General Requirements for the amount(s) shown on the accompanying bid sheet(s). By signing below, Bidder has read the entire document and agreed to the terms therein.

Signer's Name: _____ Firm Name: _____
(Please Print or Type)

Address: _____

Signature of Person Authorized to Sign Bid _____ City, State, Zip Code: _____

Email Address: _____ Telephone No.: _____

Fax No.: _____

Please complete the following:

Prompt Payment Discount: _____% _____days. (If no discount is offered, Net 30 will apply.)

Please check the following blanks which apply to your company:

Ownership of firm (51% or more):

___ Non-minority ___ Hispanic ___ African-American ___ Other Minority (specify) _____

___ Female Owned ___ Handicapped Owned ___ Small Business (less than \$1 million annual receipts or 100 employees)

Indicate Status: ___ Partnership ___ Corporation ___ Sole Proprietorship ___ Other (specify) _____

Tax Identification Number: _____

To report suspected ethics violations impacting the San Antonio Water System, please call 1-800-687-1918.

***** This Addendum 2 is issued for Bid 19-19112A.

1. Answer question(s).
2. Include Year 5 for Warranty and Service Support.

IT IS **NECESSARY** TO RETURN THIS ADDENDUM 2 AS PART OF YOUR BID SUBMISSION

.....

1.This question pertains to item number seven on the price schedule list on page 20 of the bid document. Does the item “Warranty and service support” for the 1st year begin one year after installation or at the time of installation?

Response: The first year of the warranty is included with the purchase of the instrument and begins after the installation is complete.

2.This question pertains to warranty and service support options. The bid states under item two of the technical specifications that “bidder shall provide a detail for warranty cost and service support agreement for a period of two (2) years to cover the initial purchase and three (3) additional years”. However, the pricing sheet only allows for the listing of years one through four and does not include the pricing for year five. Please advise on how you would like the pricing sheet to be filled out.

Response: Return Addendum 2 with new price schedule. Include page 1 of Addendum 2 with signature.

3. Is the Meinsberg (Xylem) pH Transmitter MV 4010 an approved equal pH and Temperature Meter for the system?

Response: Yes. As long as it meets the minimum requirements as outlined in Item 1-C of the Bid.

4.Is the Meinsberg (Xylem) pH Combination Electrode EGS 150 an approved equal pH and Temperature Probe for the system?

Response: Yes. As long as it meets the minimum requirements as outlined in Item 1-D of the Bid.

5.Is YSI Optical BOD Probe SEAL Part number 300-310018 an approved equal BOD Probe?

Response: As long as it meets the minimum requirements as outlined in Item 1-A of the Bid.

6.Is the SEAL Automated BOD analyzer pre-approved to offer a bid response?

Response: Yes

7.Per Addendum 1, am I correct that the system is not required to aquatically dose the sample into the BOD bottle after the pH of the sample is adjusted?

Response: Correct, it is not required.

8.Per Addendum 1, am I correct that the system is not required to cap and de-cap the BOD bottles?

Response: Correct, it is not required.

9. Is SEAL able to bid more than one system? One that meets the required technical specifications and another that will include all the highly desired functions requested by SAWS that will allow complete stand-alone operation and complete automation?

Response: Yes, vendors may bid on more than one system.

ESTIMATED DELIVERY

Delivery will be completed within _____ calendar days after receipt of order. (if applicable)

PRICE SCHEDULE

| Item | Description | QTY | Unit Price | UOM | Extended Price |
|------|--|-----|------------|-----|----------------|
| 1 | <p>BOD Automated Analyzer, to include: <i>Probes (4)</i> Manufacturer: <u>YSI</u> Description: <u>Pro OBOD Optical BOD probe, US</u> Part No.: <u>626400</u></p> <p>OR APPROVED EQUAL</p> <p>Manufacturer: _____ Description: _____ Part No.: _____</p> | 1 | \$ | EA | \$ |
| 2 | <p><i>Meters (2)</i> Manufacturer: <u>YSI</u> Description: <u>ProSolo ODO Optical Dissolved Oxygen Meter</u> Part No.: <u>626650</u></p> <p>OR APPROVED EQUAL</p> <p>Manufacturer: _____ Description: _____ Part No.: _____</p> | 2 | \$ | EA | \$ |
| 3 | <p><i>pH and Temperature Meter (1)</i> Manufacturer: <u>YSI</u> Description: <u>MultiLab 4010-1W Meter</u> Part No.: <u>1FD35OY</u></p> <p>OR APPROVED EQUAL</p> <p>Manufacturer: _____ Description: _____ Part No.: _____</p> | 1 | \$ | EA | \$ |
| 4 | <p><i>pH and Temperature Probe (1)</i> Manufacturer: <u>YSI</u> Description: <u>MultiLab IDS 4110 pH & Temperature Sensor</u> Part No.: <u>103740Y</u></p> <p>OR APPROVED EQUAL</p> | 1 | \$ | EA | \$ |

| | | | | | |
|----|--|----------|----|----|----|
| | Manufacturer: _____ Description: _____ Part No.: _____ | | | | |
| 5 | Data Management System (includes CPU, monitor, keyboard, mouse, printer and all hardware necessary to communicate with the BOD Analyzer) Installation | 1 | \$ | LS | \$ |
| 6 | On-Site Training After Installation and Full Completion of Manufacturer Check Out Process (4 days) Travel: arrangements shall be in accordance to the SAWS Reimbursable Expense Policy. | 1 | \$ | LS | \$ |
| 7 | Warranty and Service Support: This shall include the BOD Automated Analyzer and Data Management System | 1st year | \$ | EA | \$ |
| 8 | Warranty and Service Support: This shall include the BOD Automated Analyzer and Data Management System | 2nd year | \$ | EA | \$ |
| 9 | Warranty and Service Support: This shall include the BOD Automated Analyzer and Data Management System | 3rd year | \$ | EA | \$ |
| 10 | Warranty and Service Support: This shall include the BOD Automated Analyzer and Data Management System | 4th year | \$ | EA | \$ |
| 11 | Warranty and Service Support: This shall include the BOD Automated Analyzer and Data Management System | 5th year | \$ | EA | \$ |
| | TOTAL | | | | \$ |

IT IS **NECESSARY** TO RETURN THIS ADDENDUM 2 AS PART OF YOUR BID SUBMISSION

All other terms and conditions of the original bid remain unchanged.